



BSI Standards Publication

Wheelchair seating

Part 12: Apparatus and method
for cushion envelopment testing

National foreword

This Published Document is the UK implementation of ISO/TS 16840-12:2015.

The UK participation in its preparation was entrusted by Technical Committee CH/173, Assistive products for persons with disability, to Subcommittee CH/173/1, Wheelchairs.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2015.

Published by BSI Standards Limited 2015

ISBN 978 0 580 90009 9

ICS 11.180.10

Compliance with a British Standard cannot confer immunity from legal obligations.

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 31 March 2015.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

Wheelchair seating —

Part 12:

Apparatus and method for cushion envelopment testing

Sièges de fauteuils roulants —

Partie 12: Appareillage et méthode d'essai de l'enveloppement du coussin





COPYRIGHT PROTECTED DOCUMENT

© ISO 2015

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Indenter construction	2
4.1 Materials.....	2
4.2 Tolerances and finishes.....	2
4.3 Indenter elements.....	3
4.3.1 Trochanter bar.....	3
4.3.2 Trochanter buttons.....	3
4.3.3 Bulbous indenter half.....	4
4.4 Indenter assembly.....	5
4.5 Sensor locations.....	6
4.5.1 General.....	6
4.5.2 Elevation ₁	6
4.5.3 Elevation ₂	7
4.5.4 Elevation ₃	7
4.5.5 Elevation ₄	7
5 Instrumentation hardware	8
5.1 Electronics.....	8
5.2 Loading equipment.....	8
6 Preparation of test cushion	8
6.1 Cushion adjustment.....	8
6.2 Preconditioning the cushion to the test environment.....	8
7 Envelopment test method	8
7.1 Rationale.....	8
7.2 Method.....	8
7.2.1 Setting the datum height for the indenter.....	8
7.2.2 Pre-conditioning the cushion.....	8
7.2.3 Measuring the seat cushion thickness.....	9
7.2.4 Load application and data collection.....	9
7.3 Calculations.....	10
8 Test report	11
Annex A (normative) Method adaptations for cushions that utilize offloading	13
Bibliography	14

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT), see the following URL: [Foreword — Supplementary information](#).

The committee responsible for this document is ISO/TC 173, *Assistive products for persons with disability*, Subcommittee SC 1, *Wheelchairs*.

ISO 16840 consists of the following parts, under the general title *Wheelchair seating*:

- *Part 1: Vocabulary, reference axis convention and measures for body segments, posture and postural support surfaces*
- *Part 2: Determination of physical and mechanical characteristics of devices intended to manage tissue integrity — Seat cushions*
- *Part 3: Determination of static, impact and repetitive load strengths for postural support devices*
- *Part 4: Seating systems for use in motor vehicles*
- *Part 6: Simulated use and determination of the changes in properties — Seat cushions*
- *Part 10: Resistance to ignition of non-integrated seat and back support cushions — Requirements and test methods*
- *Part 11: Determination of perspiration dissipation characteristics of seat cushions intended to manage tissue integrity* [Technical Specification]
- *Part 12: Apparatus and method for cushion envelopment testing* [Technical Specification]

The following parts are under preparation:

- *Part 9: Clinical interface pressure mapping guidelines for seating* [Technical Report]

Future parts dealing with methods for determining heat and water vapour characteristics and clinical guideline for the measurement of postural support surfaces and body segments are planned.

Introduction

This part of ISO 16840 provides details of test equipment and a method for the measurement of “performance” of a wheelchair cushion intended to use immersion and envelopment to reduce local areas of pressure (by effectively supporting more tissue). The primary elements that represent the basic cushioning effect of a cushion are immersion and envelopment. Immersion into the cushion (the depth that a body penetrates into the surface) and the envelopment of the body (the intimacy of the cushion to the body) combine to define the potential cushioning performance of the cushion. In this test, the distribution of forces across the surface of the indenter is evaluated with multiple indenter sizes and masses. The accommodation of the cushion to the changes in indenter size and mass are representative of the changes in size and mass of the occupant that can occur in the life of a wheelchair user or between different users.

Issues related to the use of devices that measure the forces between the body and a support surface (i.e. pressure mapping systems) have led to the use of an instrumented indenter, which has fewer sensors, but sensors which are high quality and exhibit repeatability, accuracy, and thus reliability.

Wheelchair seating —

Part 12:

Apparatus and method for cushion envelopment testing

1 Scope

This part of ISO 16840 specifies apparatus, test methods, and disclosure requirements for characterization of wheelchair seat cushion immersion and envelopment properties using instrumented indenters to characterize the interface pressure of each indenter and the test cushion by measuring the cushioning effects of immersion and envelopment. This part of ISO 16840 can be considered to expand the characterization of products intended to manage tissue integrity (ISO 16840-2) and provide a standardized indenter for other wheelchair seating tests. It does not provide information specific to cushion performance for a particular individual user.

This part of ISO 16840 includes a method that is specific to 220 mm and 255 mm indenters. Dimensions and loads are provided for the 380 mm indenter to allow for extension of the methods for bariatric applications.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO/IEC Guide 98-3, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

ISO 1302, *Geometrical Product Specifications (GPS) — Indication of surface texture in technical product documentation*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

adjustable cushion

cushion in which adjustments can be made in a reversible fashion

Note 1 to entry: See [3.2](#).

3.2

cushion adjustment

modification to the volume, size, hardness, pressure, or other physical characteristics of the cushion surface to accommodate the applied load and the indenter size in accordance with the manufacturer's instructions

3.3

base points

lowest two points of indenter when ready for use

Note 1 to entry: See [Figures 6](#) and [7](#).

Note 2 to entry: The base points correspond with the ischial tuberosities on the human pelvis.