



*NSF International Standard /
American National Standard*

NSF/ANSI 49 - 2016

**Biosafety Cabinetry: Design,
Construction, Performance, and
Field Certification**



NSF International, an independent, not-for-profit, non-governmental organization, is dedicated to being the leading global provider of public health and safety-based risk management solutions while serving the interests of all stakeholders.

*This Standard is subject to revision.
Contact NSF to confirm this revision is current.*

Users of this Standard may request clarifications and interpretations, or propose revisions by contacting:

Chair, Joint Committee on Biosafety Cabinetry
c/o NSF International
789 North Dixboro Road, P.O. Box 130140
Ann Arbor, Michigan 48113-0140 USA
Phone: (734) 769-8010 Telex: 753215 NSF INTL
FAX: (734) 769-0109
E-mail: info@nsf.org
Web: <http://www.nsf.org>

NSF International Standard/
American National Standard
for Biosafety Cabinetry —

**Biosafety Cabinetry: Design,
Construction, Performance, and
Field Certification**

Standard Developer

NSF International

NSF International

Designated an ANSI Standard

June 24, 2016

American National Standards Institute

Prepared by
The NSF Joint Committee on Biosafety Cabinetry

Recommended for Adoption by
The NSF Council of Public Health Consultants

Adopted by
The NSF Board of Trustees
June 1976

Revised May 1983
Revised June 1987
Revised May 1992
Revised March 2002
Addendum November 2002
Revised February 2004
Revised September 2004
Addendum October 2004
Addendum March 2005
Revised July 2007
Revised October 2008
Revised June 2009
Revised September 2010
Revised November 2010

Revised November 2011
Revised July 2012
Revised February 2015
Revised March 2017

Published by

NSF International
PO Box 130140, Ann Arbor, Michigan 48113-0140, USA

For ordering copies or for making inquiries with regard to this Standard, please reference the designation "NSF/ANSI 49 – 2016."

Copyright 2017 NSF International
Previous editions © 2014, 2012, 2011, 2010, 2009, 2008, 2007, 2004, 2002, 1992, 1987, 1983, 1976

Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from NSF International.

Printed in the United States of America.

Disclaimers¹

NSF International (NSF), in performing its functions in accordance with its objectives, does not assume or undertake to discharge any responsibility of the manufacturer or any other party. The opinions and findings of NSF represent its professional judgment. NSF shall not be responsible to anyone for the use of or reliance upon this Standard by anyone. NSF shall not incur any obligation or liability for damages, including consequential damages, arising out of or in connection with the use, interpretation of, or reliance upon this Standard.

NSF Standards provide basic criteria to promote sanitation and protection of the public health. Provisions for mechanical and electrical safety have not been included in this Standard because governmental agencies or other national standards-setting organizations provide safety requirements.

Participation in NSF's Standards development activities by regulatory agency representatives (federal, local, state) shall not constitute their agency's endorsement of NSF or any of its Standards.

Preference is given to the use of performance criteria measurable by examination or testing in NSF Standards development when such performance criteria may reasonably be used in lieu of design, materials, or construction criteria.

The illustrations, if provided, are intended to assist in understanding their adjacent standard requirements. However, the illustrations may not include **all** requirements for a specific product or unit, nor do they show the only method of fabricating such arrangements. Such partial drawings shall not be used to justify improper or incomplete design and construction.

Unless otherwise referenced, the annexes are not considered an integral part of NSF Standards. The annexes are provided as general guidelines to the manufacturer, regulatory agency, user, or certifying organization.

¹ The information contained in this Disclaimer is not part of this American National Standard (ANS) and has not been processed in accordance with ANSI's requirements for an ANS. As such, this Disclaimer may contain material that has not been subjected to public review or a consensus process. In addition, it does not contain requirements necessary for conformance to the Standard.

This page is intentionally left blank.

Contents

1	General	1
1.1	Scope	1
1.2	Minimum requirements.....	1
1.3	Variations in design and construction	1
2	Normative references	1
3	Definitions	3
4	Materials	11
4.1	General.....	11
4.2	Interior work surfaces.....	11
4.3	Exposed interior surfaces.....	11
4.4	Other interior and exterior surfaces	11
4.5	Materials and finishes	11
5	Design and construction	13
5.1	General.....	13
5.2	Cleanability.....	13
5.3	Decontamination	13
5.4	Canopy connect exhaust.....	13
5.5	Direct connect exhaust.....	14
5.6	Duct and plenum design	14
5.7	Internal corners and angles	14
5.8	External corners and angles	14
5.9	Joints and seams	14
5.10	Fastening methods	15
5.11	Welds.....	15
5.12	Solder	15
5.13	Removable panels.....	15
5.14	Stability	15
5.15	Provision for mounting.....	15
5.16	Legs and feet.....	16
5.17	Reinforcing and framing	16
5.18	Fixed panels	16
5.19	Doors and covers	16
5.20	Louvers and openings	17
5.21	Tracks and guides	17
5.22	Filters.....	17
5.24	Stopcocks and service outlets	18
5.25	Alarms.....	18
5.26	Electrical components	19
5.27	Lighting	20
5.28	Gauges	20
5.29	Drain spillage trough.....	20
5.30	Diffuser placement.....	20
5.31	Work area components placement.....	20
5.32	Data plate(s)	20
5.33	Routine maintenance adjustment fixtures	21
6	Performance	28
6.1	General.....	28
6.2	Pressure decay / soap bubble / tracer gas leak.....	28
6.3	HEPA/ULPA filter leak.....	28
6.4	Noise level.....	28
6.5	Lighting intensity	28
6.6	Vibration	28
6.7	Personnel, product, and cross-contamination protection.....	28
6.8	Stability.....	29
6.9	Downflow and inflow velocities	29
6.10	Inflow velocity	30
6.11	Airflow smoke patterns	30

6.12	Drain spillage trough leakage	31
6.13	Motor/blower performance.....	31
6.14	Electrical safety	31
6.15	Performance data	31
6.16	Record maintenance	31
6.17	Air velocity stability	31
Annex A	33
Annex B	79
Annex C	83
Annex D	87
Annex E	89
Annex F	129
Annex G	147
Annex H	165
Annex I	169
Annex J	173
Annex K	177

Foreword²

The purpose of this Standard is to establish minimum requirements for materials, design, construction, and performance of Biosafety Cabinetry that are designed to protect personnel, product, and the environment. This Standard details requirements for performance testing as well as field certification testing.

This edition of the Standard (NSF/ANSI 49-2016) includes the following revisions:

Issue 45

Changes were made to Annex G covering the addition of Vaporized Hydrogen Peroxide as a decontamination agent in Biosafety Cabinetry, as well as clarifying the use of Chlorine Dioxide .

Issue 56

Language in section 6.14 was added regarding certification to IEC 61010-1 or a national standard based on it.

Issue 73

The newly designated C1 BSC was incorporated into the Standard.

Issue 76

Updated language was added referencing audible and visual alarms.

Issue 78

Metric conversions throughout the Standard were reviewed and updated.

Issue 79

Decontamination procedures in Annex G were updated.

Issue 81

This issue updated Annex E.

Issue 86

Updates were made in Annex A to the incubation time and temperature prior to the micro check.

Issue 88

Language in section 5.32 regarding cabinet height and width was removed.

Issue 90

Redundancies about sliding sash alarms in subsections 5.19.4 and 5.25.1 were removed.

Issue 96

Language involving preparation of the spore suspension of *Bacillus atrophaeus* 9372 in Annex A was updated.

Issue 99

The soap bubble leak test in Annex A was revised.

This Standard was developed by the NSF Joint Committee on Biosafety Cabinetry using the consensus process described by the American National Standards Institute.

² The information contained in this Foreword is not part of this American National Standard (ANS) and has not been processed in accordance with ANSI's requirements for an ANS. As such, this Foreword may contain material that has not been subjected to public review or a consensus process. In addition, it does not contain requirements necessary for conformance to the Standard.

Suggestions for improvement of this Standard are welcome. This Standard is maintained on a Continuous Maintenance schedule and can be opened for comment at any time. Comments should be sent to Chair, Joint Committee on Biosafety Cabinetry at standards@nsf.org, or c/o NSF International, Standards Department, PO Box 130140, Ann Arbor, Michigan 48113-0140, USA.

NSF/ANSI Standard
for Biosafety Cabinetry —

Biosafety Cabinetry: Design, Construction, Performance, and Field Certification

1 General

1.1 Scope

This Standard applies to Class II (laminar flow) biosafety cabinetry designed to minimize hazards inherent in work with agents assigned to biosafety levels 1, 2, 3, or 4. It also defines the tests that shall be passed by such cabinetry to meet this Standard. This Standard includes basic requirements for the design, construction, and performance of biosafety cabinets that are intended to provide personnel, product, and environmental protection; reliable operation; durability and structural stability; cleanability; limitations on noise level; illumination; vibration; and motor/blower performance.

1.2 Minimum requirements

Cabinets qualifying under this Standard shall have passed all of the designated tests. Units with component parts covered under existing NSF standards or criteria shall conform to those applicable requirements.

1.3 Variations in design and construction

Cabinetry varying in design, construction, or installation of accessory equipment may qualify under this Standard, if appropriate tests and investigations indicate that the equipment is durable and reliable, can be cleaned and decontaminated, and performs in conformance to this Standard. Such equipment shall meet the requirements for materials and finishes in this Standard.

Major modifications require appropriate tests for conformance. Major modifications include, but are not limited to, changes in the following: location or capacity or quantity or all three of blower/motor(s); size or design or both of air plenums; position of High Efficiency Particulate Air/Ultra Low Penetrating Air (HEPA/ULPA) filters; position or redesign of work surface; work area intake and exhaust air grilles; window placement or design; access opening size; location and size of exhaust port; and built-in accessory equipment (centrifuges, ultraviolet lighting, supports for intravenous drug container, arm rests, etc.). Relocation of utility service equipment (electrical outlets, petcocks, etc.) is not considered a major modification if other provisions of this Standard are not compromised.

2 Normative references

The following documents contain requirements that, by reference in this text, constitute requirements of this Standard. At the time of publication, the indicated editions were valid. All documents are subject to revision, and parties are encouraged to investigate the possibility of applying the most recent editions of the documents indicated below.