



*NSF International Standard /
American National Standard*

NSF/ANSI 49 - 2018

**Biosafety Cabinetry: Design,
Construction, Performance,
and Field Certification**



NSF International, an independent, not-for-profit, non-governmental organization, is dedicated to being the leading global provider of public health and safety-based risk management solutions while serving the interests of all stakeholders.

This Standard is subject to revision.
Contact NSF to confirm this revision is current.

Users of this Standard may request clarifications and interpretations, or propose revisions by contacting:

Chair, Joint Committee on Biosafety Cabinetry
c/o NSF International
789 North Dixboro Road, PO Box 130140
Ann Arbor, Michigan 48113-0140 USA
Phone: (734) 769-8010 Telex: 753215 NSF INTL
FAX: (734) 769-0109
E-mail: info@nsf.org
Web: www.nsf.org

NSF International Standard /
American National Standard
for Biosafety Cabinetry –

**Biosafety Cabinetry: Design,
Construction, Performance,
and Field Certification**

Standard Developer

NSF International

Designated as an ANSI Standard

March 5, 2018

American National Standards Institute

Prepared by
The NSF Joint Committee on Biosafety Cabinetry

Recommended for adoption by
The NSF Council of Public Health Consultants

Adopted by
The NSF Board of Trustees
June 1976

Revised May 1983
Revised March 2002
Revised September 2004
Revised July 2007
Revised September 2010
Revised July 2012
Revised January 2019

Revised June 1987
Addendum November 2002
Addendum October 2004
Revised October 2008
Revised November 2010
Revised February 2015

Revised May 1992
Revised February 2004
Addendum March 2005
Revised June 2009
Revised November 2011
Revised March 2017

Published by

NSF International
PO Box 130140, Ann Arbor, Michigan 48113-0140, USA

For ordering copies or for making inquiries with regard to this Standard, please reference the designation "NSF/ANSI 49 – 2018."

Copyright 2019 NSF International
Previous editions © 2017, 2014, 2012, 2011, 2010, 2009, 2008, 2007, 2004, 2002, 1992, 1987, 1983, 1976

Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from NSF International.

Printed in the United States of America.

Disclaimers¹

NSF International (NSF), in performing its functions in accordance with its objectives, does not assume or undertake to discharge any responsibility of the manufacturer or any other party. The opinions and findings of NSF represent its professional judgment. NSF shall not be responsible to anyone for the use of or reliance upon this Standard by anyone. NSF shall not incur any obligation or liability for damages, including consequential damages, arising out of or in connection with the use, interpretation of, or reliance upon this Standard.

NSF Standards provide basic criteria to promote sanitation and protection of the public health. Provisions for mechanical and electrical safety have not been included in this Standard because governmental agencies or other national standards-setting organizations provide safety requirements.

Participation in NSF's Standards development activities by regulatory agency representatives (federal, local, state) shall not constitute their agency's endorsement of NSF or any of its Standards.

Preference is given to the use of performance criteria measurable by examination or testing in NSF Standards development when such performance criteria may reasonably be used in lieu of design, materials, or construction criteria.

The illustrations, if provided, are intended to assist in understanding their adjacent standard requirements. However, the illustrations may not include all requirements for a specific product or unit, nor do they show the only method of fabricating such arrangements. Such partial drawings shall not be used to justify improper or incomplete design and construction.

Unless otherwise referenced, the Annexes are not considered an integral part of NSF Standards. The Annexes are provided as general guidelines to the manufacturer, regulatory agency, user, or certifying organization.

¹ The information contained in this Disclaimer is not part of this American National Standard (ANS) and has not been processed in accordance with ANSI's requirements for an ANS. As such, this Disclaimer may contain material that has not been subjected to public review or a consensus process. In addition, it does not contain requirements necessary for conformance to the Standard.

Contents

1	General	1
1.1	Scope	1
1.2	Minimum requirements.....	1
1.3	Variations in design and construction	1
2	Normative references	1
3	Definitions	3
4	Materials	11
4.1	General.....	11
4.2	Interior work surfaces.....	11
4.3	Exposed interior surfaces.....	11
4.4	Other interior and exterior surfaces.....	11
4.5	Materials and finishes	11
5	Design and construction	13
5.1	General.....	13
5.2	Cleanability.....	13
5.3	Decontamination	13
5.4	Canopy exhaust connection.....	13
5.5	Direct exhaust connection.....	14
5.6	Duct and plenum design	14
5.7	Internal corners and angles.....	14
5.8	External corners and angles	14
5.9	Joints and seams	14
5.10	Fastening methods.....	15
5.11	Welds	15
5.12	Solder	15
5.13	Removable panels	15
5.14	Stability.....	15
5.15	Provision for mounting	15
5.16	Legs and feet.....	16
5.17	Reinforcing and framing	16
5.18	Fixed panels	16
5.19	Doors and covers	16
5.20	Louvers and openings	17
5.21	Tracks and guides.....	17
5.22	Filters.....	17
5.23	Gaskets and sealants.....	18
5.24	Stopcocks and service outlets.....	18
5.25	Alarms	18
5.26	Electrical components	20
5.27	Lighting.....	20
5.28	Gauges.....	20
5.29	Drain spillage trough	21
5.30	Diffuser placement	21
5.31	Work area components placement	21
5.32	Data plate(s).....	21
5.33	Routine maintenance adjustment fixtures.....	21
6	Performance	28

6.1	General.....	28
6.2	Pressure decay / soap bubble / tracer gas leak.....	28
6.3	HEPA/ULPA filter leak.....	28
6.4	Noise level.....	28
6.5	Lighting intensity	28
6.6	Vibration	28
6.7	Personnel, product, and cross-contamination protection.....	28
6.8	Stability.....	29
6.9	Downflow and inflow velocities.....	29
6.10	Inflow velocity.....	30
6.11	Airflow smoke patterns.....	30
6.12	Drain spillage trough leakage.....	30
6.13	Motor / blower performance	30
6.14	Electrical safety	31
6.15	Performance data.....	31
6.16	Record maintenance	31
6.17	Air velocity stability.....	31
Annex A	Performance tests.....	33
A.1	Pressure decay or soap bubble	33
A.2	HEPA/ULPA filter leak test.....	34
A.3	Noise level test.....	36
A.4	Lighting intensity test.....	37
A.5	Vibration test	37
A.6	Personnel, product, and cross-contamination protection (biological) tests	38
A.7	Stability tests	49
A.8	Downflow velocity.....	50
A.9	Inflow velocity (face velocity) test.....	52
A.10	Airflow smoke patterns test.....	55
A.11	Drain spillage trough leakage test.....	56
A.12	Motor or blower performance	57
A.13	Cabinet airflow stability	58
A.14	Canopy connection test.....	60
Annex B	83
B.1	Method to verify fitness for use of potential direct inflow measurement devices.....	83
Annex C	Nebulizer selection and calibration.....	85
C.1	Selection.....	85
C.2	Calibration	85
Annex D	Evaluation of chemical resistance and abrasion resistance of surfaces	89
D.1	Chemical resistance.....	89
D.2	Abrasion resistance.....	89
Annex E	Biosafety cabinet selection, installation, use, lifespan, and decommissioning.....	91
E.1	Institutional safety consultation	91
E.2	Risk assessment procedure.....	91
E.3	BSC Class and Type selection.....	96
E.4	Site review before BSC purchase	104
E.5	BSC arrival and certification.....	106

E.6	Cleaning and disinfection of BSC work area.....	107
E.7	BSC use practices and procedures.....	109
E.8	Moving a permanently installed BSC	113
E.9	BSC lifespan.....	113
E.10	Decommissioning process	113
E.11	Definitions.....	115
Annex F	Field tests.....	133
F.1	Field certification preconditions and intervals	133
F.2	Downflow velocity.....	134
F.3	Inflow velocity (face velocity) test.....	135
F.4	Airflow smoke patterns test.....	139
F.5	HEPA/ULPA filter leak test.....	141
F.6	Pressure decay / soap bubble.....	143
F.7	Site installation assessment tests	144
F.8	Electrical leakage and ground circuit resistance and polarity tests	147
F.9	Lighting intensity test.....	147
F.10	Vibration test	148
F.11	Noise level tests	149
F.12	Record of field certification.....	150
Annex G	Field tests.....	151
G.1	Recommended biosafety cabinet decontamination procedure.....	151
G.2	Recommended HEPA/ULPA filter disposal procedures	167
Annex H	Recommended materials, finishes, and construction	169
H.1	Sheet metal and finishes.....	169
H.2	Glass	169
H.3	HEPA/ULPA filter gasket materials.....	169
H.4	HEPA/ULPA filter case – Type IC.....	170
H.5	Specifications	170
H.6	Sealants	170
H.7	Fans	170
H.8	Components and wiring	171
Annex I	Reference standards and specifications pertinent to Class II biosafety cabinetry	173
I.1	Miscellaneous publications	173
I.2	Federal specifications	174
I.3	Federal standards	174
I.4	Military specifications	175
Annex J	177
J.1	Helium leak test.....	177
J.2	Sulfur hexafluoride (SF ₆) leak test	178
Annex K	Protocol for the validation of alternate biosafety cabinet decontaminating methods and agents....	181
K.1	Introduction.....	181
K.2	Protocol	181

Foreword²

The purpose of this Standard is to establish minimum requirements for materials, design, construction, and performance of Biosafety Cabinetry that are designed to protect personnel, product, and the environment. This Standard details requirements for performance testing as well as field certification testing.

This edition includes the following revisions:

Issue 47

Revised and added new language in Annex A regarding the cross contamination test procedure.

Issue 59

Revised and added new language for the airflow alarm requirement language for all cabinet types.

Issue 77

Added language to Annex A regarding confirmation tests when there is a test failure.

Issue 105

Added new language in Annex E regarding Risk assessment of biosafety cabinets exhaust system pressurization in the event of an exhaust system failure.

Issue 108

Added new language in Annex F regarding exhaust airflow alarms.

Issue 109

Revised and added new language referring to the term Certification throughout this Standard.

Issue 110

Revised language in Section 1.3 regarding the identification of major software changes to biosafety cabinetss.

Issue 111

Added new and revised language to Section 3 regarding the terms visible and viewable, and added provisions for the optional use of digital data plates to Section 5.

Issue 112

Revised and added new language in Annex G regarding the generation and dispersion of decontamination gas.

² The information contained in this Foreword is not part of this American National Standard (ANS) and has not been processed in accordance with ANSI's requirements for an ANS. As such, this Foreword may contain material that has not been subjected to public review or a consensus process. In addition, it does not contain requirements necessary for conformance to the Standard.

Issue 115

Added new language in Section 6.10 regarding inflow requirements for Type C1 Cabinets.

Issue 117

Updates the language throughout this Standard regarding the use of the term “and/or”.

Issue 118

Revised language in Section 5.25.3 and Annex F regarding the exhaust alarm in Type B biosafety cabinets.

Issue 121

Revised language in Section 5.26.2 regarding electrical wiring.

Issue 125

Revised language in Section 5.26.2 regarding the term ‘running power’.

This Standard was developed by the NSF Joint Committee on Biosafety Cabinetry using the consensus process described by the American National Standards Institute.

Suggestions for improvement of this Standard are welcome. This Standard is maintained on a Continuous Maintenance schedule and can be opened for comment at any time. Comments should be sent to Chair, Joint Committee on Biosafety Cabinetry at standards@nsf.org, or c/o NSF International, Standards Department, PO Box 130140, Ann Arbor, Michigan 48113-0140, USA.

NSF/ANSI Standard
for Biosafety Cabinetry —

Biosafety Cabinetry: Design, Construction, Performance, and Field Certification

1 General

1.1 Scope

This Standard applies to Class II (laminar flow) biosafety cabinetry designed to minimize hazards inherent in work with agents assigned to Biosafety Levels 1, 2, 3, or 4. It also defines the tests that shall be passed by such cabinetry to meet this Standard. This Standard includes basic requirements for the design, construction, and performance of biosafety cabinets (BSCs) that are intended to provide personnel, product, and environmental protection; reliable operation; durability and structural stability; cleanability; limitations on noise level; illumination; vibration; and motor / blower performance.

1.2 Minimum requirements

Cabinets qualifying under this Standard shall have passed all of the designated tests. Units with component parts covered under existing NSF Standards or Criteria shall conform to those applicable requirements.

1.3 Variations in design and construction

Cabinetry varying in design, construction, or installation of accessory equipment may qualify under this Standard, if appropriate tests and investigations indicate that the equipment is durable and reliable, can be cleaned and decontaminated, and performs in conformance to this Standard. Such equipment shall meet the requirements for materials and finishes in this Standard.

Major modifications require appropriate tests for conformance. Major modifications include, but are not limited to any of the following changes to the blower / motor(s): location, capacity, quantity, or automatic airflow adjustment; size, or design, or both, of air plenums; position of high efficiency particulate air / ultra low penetrating air (HEPA/ULPA) filters; position or redesign of work surface; work area intake and exhaust air grilles; window placement or design; access opening size; location and size of exhaust port; the visibility or audibility of the safety signaling; and built-in accessory equipment (centrifuges, ultraviolet (UV) lighting, supports for intravenous drug container, arm rests, etc.). Major modifications also include changes affecting the safe use of the cabinet including the ability to see, hear, and understand the required alarms. Relocation of utility service equipment (electrical outlets, petcocks, etc.), the visual appearance of the cabinet, or user interface(s), are not considered a major modification if other provisions of this Standard are not compromised.

2 Normative references

The following documents contain requirements that, by reference in this text, constitute requirements of this Standard. At the time of publication, the indicated editions were valid. All documents are subject to revision, and parties are encouraged to investigate the possibility of applying the most recent editions of the documents indicated below.