

Australian Standard™

**Medical devices—Quality management
systems—Requirements for regulatory
purposes**

This Australian Standard was prepared by Committee HE-028, Quality Management and Corresponding General Aspects for Medical Devices. It was approved on behalf of the Council of Standards Australia on 24 Dec 2003 and published on 31 December 2003.

The following are represented on Committee HE-028:

- Australian Chamber of Commerce and Industry
- Australian Dental Association
- Australian Dental Industry Association Inc
- Australian Healthcare Association
- Australian Organisation for Quality
- Australian Private Hospitals Association
- Certification Bodies (Australia)
- Department of Human Services (South Australia)
- Department of Human Services (Victoria)
- Health Department of Western Australia
- Medical Industry Association of Australia Inc
- New South Wales Health Department
- Quality Society of Australasia
- Queensland Health
- The Institute of Internal Auditors
- Therapeutic Goods Administration

Keeping Standards up-to-date

Standards are living documents which reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about Standards can be found by visiting the Standards Web Shop at www.standards.com.au and looking up the relevant Standard in the on-line catalogue.

Alternatively, the printed Catalogue provides information current at 1 January each year, and the monthly magazine, *The Global Standard*, has a full listing of revisions and amendments published each month.

Australian Standards™ and other products and services developed by Standards Australia are published and distributed under contract by SAI Global, which operates the Standards Web Shop.

We also welcome suggestions for improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at mail@standards.org.au, or write to the Chief Executive, Standards Australia International Ltd, GPO Box 5420, Sydney, NSW 2001.

This Standard was issued in draft form for comment as DR 03526.

Australian Standard™

Medical devices—Quality management systems—Requirements for regulatory purposes

Originated as AS ISO 13485—2002 and AS ISO 13488—2002.
AS ISO 13485—2002 and AS ISO 13488—2002 revised, amalgamated and redesignated as AS ISO 13485—2003.

COPYRIGHT

© Standards Australia International

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Published by Standards Australia International Ltd
GPO Box 5420, Sydney, NSW 2001, Australia

ISBN 0 7337 5715 4

PREFACE

This Standard was prepared by the Standards Australia Committee HE-028, Quality Management and Corresponding General Aspects for Medical Devices.

This Standard is identical with and has been reproduced from ISO 13485:2003, *Medical devices—Quality management systems—Requirements for regulatory purposes*.

This Standard cancels and replaces AS ISO 13485—2002, *Quality systems—Medical devices—Particular requirements for the application of ISO 9001*, and AS ISO 13488—2002, *Quality systems—Medical devices—Particular requirements for the application of ISO 9002*).

The objective of this Standard is to specify requirements for a quality management system where an organization needs to demonstrate its ability to provide medical devices and related services that consistently meet customer requirements and regulatory requirements applicable to medical devices and related services. The primary objective is to facilitate harmonized medical device regulatory requirements for quality management systems. As a result, it includes some particular requirements for medical devices and excludes some of the requirements of ISO 9001 that are not appropriate as regulatory requirements. Because of these exclusions, organizations whose quality management systems conform to this International Standard cannot claim conformity to ISO 9001 unless their quality management systems conform to all the requirements of ISO 9001.

The terms ‘normative’ and ‘informative’ are used to define the application of the annex to which they apply. A normative annex is an integral part of a standard, whereas an informative annex is only for information and guidance.

As this Standard is reproduced from an international Standard, the following applies:

- (a) Its number does not appear on each page of text and its identity is shown only on the cover and title page.
- (b) In the source text ‘this International Standard’ should read ‘this Australian Standard’.
- (c) A full point substitutes for a comma when referring to a decimal marker.

References to International Standards should be replaced by references to Australian or Australian/New Zealand Standards as follows:

<i>Reference to International Standard</i>		<i>Australian Standard</i>	
ISO		AS/NZS ISO	
9000	Quality management systems— Fundamentals and vocabulary	9000	Quality management systems— Fundamentals and vocabulary

CONTENTS

0	Introduction	iv
0.1	General	iv
0.2	Process approach	iv
0.3	<i>Relationship with other standards</i>	v
0.4	Compatibility with other management systems	v
1	Scope.....	1
1.1	General	1
1.2	Application.....	1
2	Normative references	2
3	Terms and definitions	2
4	Quality management system	4
4.1	General requirements	4
4.2	Documentation requirements	4
5	Management responsibility.....	6
5.1	Management commitment.....	6
5.2	Customer focus	6
5.3	Quality policy.....	6
5.4	Planning	7
5.5	Responsibility, authority and communication.....	7
5.6	Management review	8
6	Resource management.....	8
6.1	Provision of resources	8
6.2	Human resources	9
6.3	Infrastructure	9
6.4	Work environment.....	9
7	Product realization	10
7.1	Planning of product realization	10
7.2	Customer-related processes.....	10
7.3	Design and development.....	11
7.4	Purchasing.....	13
7.5	Production and service provision	14
7.6	Control of monitoring and measuring devices	17
8	Measurement, analysis and improvement.....	17
8.1	General	17
8.2	Monitoring and measurement.....	18
8.3	Control of nonconforming product	19
8.4	Analysis of data.....	19
8.5	Improvement.....	20
	Annex A (informative) Correspondence between ISO 13485:2003 and ISO 13485:1996.....	21
	Annex B (informative) Explanation of differences between ISO 13485:2003 and ISO 9001:2000	25
	Bibliography	57

INTRODUCTION

0.1 General

This International Standard specifies requirements for a quality management system that can be used by an organization for the design and development, production, installation and servicing of medical devices, and the design, development, and provision of related services.

It can also be used by internal and external parties, including certification bodies, to assess the organization's ability to meet customer and regulatory requirements.

Information marked "NOTE" is for guidance in understanding or clarifying the associated requirement.

It is emphasized that the quality management system requirements specified in this International Standard are complementary to technical requirements for products.

The adoption of a quality management system should be a strategic decision of an organization. The design and implementation of an organization's quality management system is influenced by varying needs, particular objectives, the products provided, the processes employed and the size and structure of the organization. It is not the intent of this International Standard to imply uniformity in the structure of quality management systems or uniformity of documentation.

There is a wide variety of medical devices and some of the particular requirements of this International Standard only apply to named groups of medical devices. These groups are defined in Clause 3.

0.2 Process approach

This International Standard is based on a process approach to quality management.

Any activity that receives inputs and converts them to outputs can be considered as a process.

For an organization to function effectively, it has to identify and manage numerous linked processes.

Often the output from one process directly forms the input to the next.

The application of a system of processes within an organization, together with the identification and interactions of these processes, and their management, can be referred to as the "process approach".

0.3 Relationship with other standards

0.3.1 Relationship with ISO 9001

While this is a stand-alone standard, it is based on ISO 9001.

Those clauses or subclauses that are quoted directly and unchanged from ISO 9001 are in normal font. The fact that these subclauses are presented unchanged is noted in Annex B.

Where the text of this International Standard is not identical to the text of ISO 9001, the sentence or indent containing that text as a whole is shown in italics (in blue italics for electronic versions). The nature and reasons for the text changes are noted in Annex B.

0.3.2 Relationship with ISO/TR 14969

ISO/TR 14969 is a Technical Report intended to provide guidance for the application of ISO 13485.

0.4 Compatibility with other management systems

This International Standard follows the format of ISO 9001 for the convenience of users in the medical device community.

This International Standard does not include requirements specific to other management systems, such as those particular to environmental management, occupational health and safety management, or financial management.

However, this International Standard enables an organization to align or integrate its own quality management system with related management system requirements. It is possible for an organization to adapt its existing management system(s) in order to establish a quality management system that complies with the requirements of this International Standard.

AUSTRALIAN STANDARD

Medical devices — Quality management systems — Requirements for regulatory purposes

1 Scope

1.1 General

This International Standard specifies requirements for a quality management system where an organization needs to demonstrate its ability to provide medical devices and related services that consistently meet customer requirements and regulatory requirements applicable to medical devices and related services.

The primary objective of this International Standard is to facilitate harmonized medical device regulatory requirements for quality management systems. As a result, it includes some particular requirements for medical devices and excludes some of the requirements of ISO 9001 that are not appropriate as regulatory requirements. Because of these exclusions, organizations whose quality management systems conform to this International Standard cannot claim conformity to ISO 9001 unless their quality management systems conform to all the requirements of ISO 9001 (see Annex B).

1.2 Application

All requirements of this International Standard are specific to organizations providing medical devices, regardless of the type or size of the organization.

If regulatory requirements permit exclusions of design and development controls (see 7.3), this can be used as a justification for their exclusion from the quality management system. These regulations can provide alternative arrangements that are to be addressed in the quality management system. It is the responsibility of the organization to ensure that claims of conformity with this International Standard reflect exclusion of design and development controls [see 4.2.2 a) and 7.3].

If any requirement(s) in Clause 7 of this International Standard is(are) not applicable due to the nature of the medical device(s) for which the quality management system is applied, the organization does not need to include such a requirement(s) in its quality management system [see 4.2.2 a)].

The processes required by this International Standard, which are applicable to the medical device(s), but which are not performed by the organization, are the responsibility of the organization and are accounted for in the organization's quality management system [see 4.1 a)].

In this International Standard the terms "if appropriate" and "where appropriate" are used several times. When a requirement is qualified by either of these phrases, it is deemed to be "appropriate" unless the organization can document a justification otherwise. A requirement is considered "appropriate" if it is necessary in order for

- the product to meet specified requirements, and/or*
- the organization to carry out corrective action.*